



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST PAI-1 Activity

Product Code : RK019A

Lot No. : FF0004

Expiration date : 2029-02-26

Components	Quantity	Lot	Expiration Date
tPA pre-coated plate	12x8 wells	FF0170	2029-06-25
(h) PAI-1 standard	3 vials	FE4271	2029-05-24
(h) PAI-1 Control I	1 vial	FE4272	2029-05-24
(h) PAI-1 Control II	1 vial	FF0319	2029-07-12
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FF0274	2029-07-09
F-Sample diluent	2x50 ml	FE4201	2029-05-23
Wash solution	1x50 ml	FE4202	2029-05-21
Conjugate diluent	1x25 ml	FE4203	2029-05-21
TMB	1x25 ml	250301D02	2029-03-31
Sulfuric Acid 0,45M	1x6 ml	FE3546	2029-02-26

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.00	1.94
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.011
Detection threshold	\leq 0.5 ng/ml	< 0.1 ng/ml

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.439	1.511	0.852	0.351	0.205

	Acceptance Range	MV
CI	[6.5 - 8.8]	8.0
CII	[2.9 - 4.4]	3.6

MV = Measured value

This is to certify that above product has been inspected and approved the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

27 FEV. 2026

Quality Control Manager :

P. I. CORNUEJOLS

