



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST (h) PAI-1 Antigen

Product Code : RK012A

Lot No. : FF0644 / FF0732

Expiration date : 2029-05-28

Components	Quantity	Lot	Expiration Date
Anti-(h)-PAI-1 pre-coated plate	12x8 wells	FF0849	2029-09-18
(h) PAI-1 standard	3 vials	FF0841	2029-09-13
(h) PAI-1 Control I	1 vial	FF0842	2029-09-13
(h) PAI-1 Control II	1 vial	FF0843	2029-09-13
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FF0964	2029-10-17
F-Sample diluent	2x50 ml	FF0766	2029-09-03
Wash solution	1x50 ml	FF0720	2029-08-27
Conjugate diluent	1x25 ml	FF0767	2029-09-04
TMB	1x25 ml	250904D02	2029-09-30
Sulfuric Acid 0,45M	1x6 ml	FE4263	2029-05-28

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml ≥ 1.40	1.48
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.000
Detection threshold	≤ 0.5 ng/ml	< 0.1

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.730	1.019	0.546	0.229	0.119

MV = Measured value

	Acceptance Range	MV
CI	[35.0 - 47.4]	41.5
CII	[12.5 - 18.7]	15.4

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

Quality Control Manager :

27 MAI 2026

plo S.MONEUS