HYPHEN BioMed
A Company of Sysmex Group

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : FF1356 / FF1701

Expiration date : 2029-09-30

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FF1840	2030-01-28
vWF standard	3 vials	FF1666	2030-01-02
vWF Control I	1 vial	FF1664	2030-01-02
vWF Control II	1 vial	FF1665	2030-01-02
Anti-(h)- vWF HRP immunoconjugate	3 vials	FF2024	2030-02-18
Sample diluent	2x50 ml	FF1008	2029-10-01
Wash solution	1x50 ml	FF1466	2029-12-04
Conjugate diluent	1x25 ml	FF1467	2029-12-03
TMB	1x25 ml	250904D02	2029-09-30
Sulfuric Acid 0,45M	1x6 ml	FF1007	2029-10-01

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	1.80
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.000
Detection threshold	≤ 5%	< 1%

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.105	1.537	0.972	0.455	0.252

	Acceptance Range	MV
CI	[61 - 83]	70
CII	[19 - 29]	23

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : Quality Control Approval 16 SEP. 2026

Quality Control Manager : S.MONEUS

F820-11/ZY/038A v10