

**CERTIFICATE OF ANALYSIS**

Product name:	<b>Tris Buffered Saline-tablets, pH 7.6</b>
Production date:	2025-11
Date of release:	2025-12-12
Expiration date:	2030-11
Form:	White, round tablet. 28 mm diameter
Composition:	1 tablet dissolved in 500 ml of deionized water yields: 0.15 M NaCl 0.050 M Tris-HCl Buffer

Analysis	Specification	Result
Insoluble matter	Passes filter test	Pass
Optical clarity	Clear Solution	Pass
Osmolality	362 ± 4 % (348 – 376) mOsmol/kg	Pass
pH	7.60 ± 0.10 at 25°C	Pass
Weight	8.15 g ± 2 % (7.99 g – 8.31 g)	Pass

The above material has met all quality specifications and has been reviewed by a quality representative.

*Sebastian Määttä*

2025-12-12

Quality Assurance, Sebastian Määttä

Date