



BIOPHEN™ ANTI-IIa (2 Stages Heparin Assay)

Product Ref. N°	A220005
Kit presentation	Chromogenic method, R1: ATIII (human); R2: Human Thrombin; R3: Substrate
Status	RUO

Intended use:

Two stages Anti-IIa chromogenic assay for Heparin (UFH or LMWH) activity measurement on human citrated plasma, or purified medium, with manual or automated method, according to Pharmacopeia.

Assay principle:

Heparin binding to Anti-thrombin (AT) induces a conformational change, thereby increasing AT's inactivation of factor IIa (thrombin). The residual factor IIa, not inhibited by the Heparin-AT complex, is quantified with a chromogenic substrate. The amount of pNA released (measured at 405 nm) is then dependent on the residual factor IIa. There is an inverse relationship between the concentration of heparin and color development.

Heparin + AT $\longrightarrow$ [AT Hep.]

[AT Hep.] + [FIIa (excess)] $\longrightarrow$ [FIIa-AT-Hep.] + [residual FIIa]

[residual FIIa] + Substrate $\longrightarrow$ Peptide + pNA

Key Characteristics & Advantages:

Complete solution compliant with Pharmacopeias

Easy and fast :

- Lyophilised and ready to use reagents, rapid (total assay time < 3 min with automated methods and <6 min with manual methods)
 - Easy to use on major coagulation analyzers, microplate or with basic equipment
- Method applications: Sysmex CS-series, STA-R® series

Optimized consumption with good stability

- After reconstitution: 15 days at 2-8°C, 4 days at 18-25°C, 6 months frozen
- On board: 18 hours (CS series and STA-R® series)
- Shelf life: 30 months

Extended measurement range useful to explore various contexts

- CS series: 0 – 1.25 IU/mL (UFH)
- STA-R® series: 0 – 1.25 IU/mL (UFH)


Excellent Precision with low CVs:

- CS series: Within-run < 3 % (UFH)
- STA-R® series: Within-run < 7 % (UFH)

Specific and reliable assay, reduced interferences related to sample content
(Hemoglobin, Bilirubin, Intralipids, Rivaroxaban, Apixaban, Dabigatran)

Safe and standardized method for optimal testing

- Separate packaging for buffers to be compliant with each pharmacopeia
- High quality raw materials used
- Calibrators/controls traceable to WHO International Standards





BIOPHEN™ ANTI-IIa (2 Stages Heparin Assay)

Product Ref. N°	A220005
Kit presentation	Chromogenic method, R1: ATIII (human); R2: Human Thrombin; R3: Substrate
Status	RUO

Assay Characteristics*:

Standardized method for manual or automated use, compliant with pharmacopeia

Calibration curves:



Precision (Intra-series):

CS-5100 (n = 10 determinations)

- CV% 2.3% (UFH, 0.18 IU/mL)
- CV% 1.1% (UFH, 0.47 IU/mL)

STA-R® (n = 10 determinations)

- CV% 6.6% (UFH, 0.24 IU/mL)
- CV% 4.1% (UFH, 0.50 IU/mL)

Interferences:

CS-5100: No interferences up to....

- 1000 mg/dL Hemoglobin, 60 mg/dL Bilirubin, 1000 mg/dL Intralipids, 400 ng/mL Rivaroxaban, 400 ng/mL Apixaban, 50 ng/mL Dabigatran

STA-R®: No interferences up to....

- 500 mg/dL Hemoglobin, 60 mg/dL Bilirubin, 1000 mg/dL Intralipids, 400 ng/mL Rivaroxaban, 400 ng/mL Apixaban, 400 ng/mL Dabigatran

Reagents preparation:

After reconstitution of each reagent by 1 mL of distilled water, just before use, dilute 1/5 in the appropriate buffer according to the Heparin to be assayed:

	Dilution of reagent		Buffer ⁽¹⁾	
Heparin	LMWH	UFH	LMWH	UFH
R1	1/5	1/5	AAR005L	AAR030K**
R2	1/5	1/5	AAR005L	AAR030K**
R3	1/5	1/5	AAR029K	Distilled water

Calibrators, samples and controls preparation:

Just before use, dilute 1/25 in the appropriate buffer according to the Heparin to be assayed and the pharmacopeia followed:

Heparin	Dilution	Buffer ⁽¹⁾
LMWH	1/25	AAR005L (EP) or AAR028K (USP)
UFH	1/25	AAR030K**

Volumes and incubation times are harmonized for an easier handling and automation.
Final assay concentrations are compliant with Pharmacopeias.

Kit presentation:



Kit contents: 2x1 mL (R1),
2x1 mL (R2) and 2 x1 mL (R3)

Related products:

Needed	Product name	Ref N°:
Calibrators / Controls	BIOPHEN™ UFH Control Plasma	A223101 / A224101 / A223901
	BIOPHEN™ LMWH Control Plasma	A223001 / A223801 / A224201
	BIOPHEN™ LMWH Control Low	A223701 / A224301 / A224401
	BIOPHEN™ Heparin Calibrator	A222001
	BIOPHEN™ UFH Calibrator	A222301
⁽¹⁾ Buffers	Tris-NaCl-EDTA-PEG, pH 8.40	AAR030K
	Tris-NaCl-EDTA-BSA, pH 8.40 (optional)	AAR031K
	Tris-NaCl-BSA, pH 7.40	AAR005L
	Tris-NaCl, pH 7.40	AAR028K
	Tris-EDTA-NaCl, pH 8.40	AAR029K

Related products	Product name	Ref N°:
Reagents	BIOPHEN™ Heparin LRT	A221011-RUO/A221013-RUO/A221015-RUO
	BIOPHEN™ Heparin	A221003 / A221006
	BIOPHEN™ Anti-Xa (2 Stages Heparin Assay)	A221005

* The values cited for the performance characteristics of the assay represent example of typical results and are not to be regarded as specifications for the test.

** AR031K buffer can be used. instead of AR030K buffer.