

U-CyTech BV

Yalelaan 48
3584 CM Utrecht
The Netherlands

P +31.30.253 5960

F +31.30.253 9344

INFO@ucytech.com


www.ucytech.com

Typical data Human MCP-1 ELISA kit

Catalogue no.: CT521A (5-plate format)

Validated for:

Human



Example of a representative Human MCP-1 standard curve using a 4-parameter logistic fit.

Recommended assay range: 8-512 pg/ml.

Sensitivity: 8 pg/ml.

Specificity: natural and recombinant human MCP-1.

Calibration: recombinant human MCP-1.

Type of sample: serum, plasma and culture supernatant.

PROUDLY DISTRIBUTED BY

ANIARA
ANIARA DIAGNOSTICA LLC

+1 (513) 770-1991 | +1 (866) 783-3797

7768 Service Center Drive

West Chester, OH 45069

info@aniara.com

Version 150319