

U-CyTech BV

Yalelaan 48
3584 CM Utrecht
The Netherlands
P +31.30.253 5960
F +31.30.253 9344
INFO@ucytech.com
www.ucytech.com

Data sheet Monkey IFN- γ ELISPOT antibody pair; 20-plate format

Cat. No.: CT610-20

Coating antibodies (1 vial)

Product: Monoclonal antibody to monkey interferon gamma (IFN- γ)
Isotype: Mouse IgG₁
Production: *In vitro* using serum free medium
Purification: DEAE ion exchange chromatography
Contents: Each vial contains sufficient material for coating of twenty 96-well ELISPOT plates
Buffer: Prior to lyophilization: 1.0 ml PBS + 125 mM trehalose
Application: Coating antibody in an ELISPOT system
Reconstitution: Dissolve the contents of the vial by injection of 1.0 ml distilled water into the vial and dilute 100 times in PBS. The total amount of one vial is sufficient for twenty 96-well ELISPOT plates (1920 determinations; 50 μ l/well).

Detection antibodies (1 vial)

Product: Biotinylated polyclonal antibody to monkey interferon gamma (IFN- γ)
Isotype: Rabbit Ig
Purification: Ammonium sulphate precipitation, protein A- and ligand-affinity chromatography
Labeling: With Biotin-7-NHS (N-hydroxysuccinimide)
Contents: Each vial contains sufficient material for twenty 96-well ELISPOT plates
Buffer: Prior to lyophilization: 2.0 ml PBS + 1% BSA + 125 mM trehalose
Application: Detection antibody in an ELISPOT system
Reconstitution: Dissolve the contents of the vial by injection of 2.0 ml distilled water into the vial and dilute 100 times in Dilution buffer (see Technical Data Sheet). The total amount of one vial is sufficient for twenty 96-well ELISPOT plates (1920 determinations; 100 μ l/well).

General

Specificity: Validated for detecting rhesus macaque, cynomolgus monkey, pig-tailed macaque, Japanese macaque, crested black macaque, barbary macaque, lion-tailed macaque, baboon, mandrill, African green monkey, black mangabey and Hanuman langur IFN- γ
Sterility: Membrane filtered (0.2 μ m)
Stability: The lyophilized products are stable for at least one year at 4°C (expiry date is indicated on the vials).
After reconstitution, the antibodies are stable for several months at 4°C (if kept sterile) or for minimal one year at -20°C.
References: Brown, K. *et al.* 2003. J. Immunol. 171: 6875-6882
Latta-Mahieu, M. *et al.* 2002. Hum. Gene Ther. 13: 1611-1620
Leonard, V.H. *et al.* 2010. J. Virol. 84: 3413-3420
Liang, X. *et al.* 2005. J. Virol. 79: 12321-12331
Ou, Y. *et al.* 2007. Virology 364: 291-300
Stittelaar, K.J. *et al.* 2002. Vaccine 20: 2921-2927
Zhang, Z-Q. *et al.* 2004. Virology 320: 75-84

PROUDLY DISTRIBUTED BY 
ANIARA DIAGNOSTICA LLC
+1 (513) 770-1991 | +1 (866) 783-3797
7768 Service Center Drive
West Chester, OH 45069
info@aniara.com