

Data sheet Mouse IL-4 ELISPOT antibody pair; 10-plate format

Cat. No.: CT657-10

Coating antibodies (2 vials)

Product: Monoclonal antibody to mouse interleukin 4 (IL-4)
Isotype: Rat IgG₁
Production: *In vitro* using serum free medium
Purification: Ion exchange chromatography
Contents: Each vial contains sufficient material for coating of five 96-well ELISPOT plates
Buffer: Prior to lyophilization: 0.25 ml PBS + 125 mM trehalose
Application: Coating antibody in an ELISPOT system
Reconstitution: Dissolve the contents of one vial by injection of 0.25 ml distilled water into the vial and dilute 100 times in PBS. The total amount of one vial is sufficient for five 96-well ELISPOT plates (480 determinations; 50 µl/well).

Detection antibodies (2 vials)

Product: Biotinylated monoclonal antibody to mouse interleukin 4 (IL-4)
Isotype: Rat IgG₁
Production: *In vitro* using serum free medium
Purification: Protein G-affinity chromatography
Labeling: With Biotin-7-NHS (N-hydroxysuccinimide)
Contents: Each vial contains sufficient material for five 96-well ELISPOT plates
Buffer: Prior to lyophilization: 0.5 ml PBS + 1% BSA + 125 mM trehalose
Application: Detection antibody in an ELISPOT system
Reconstitution: Dissolve the contents of one vial by injection of 0.5 ml distilled water into the vial and dilute 100 times in Dilution buffer (see Technical Data Sheet). The total amount of one vial is sufficient for five 96-well ELISPOT plates (480 determinations; 100 µl/well).

General

Specificity: Validated for detecting mouse IL-4
Sterility: Membrane filtered (0.2 µm)
Stability: The lyophilized products are stable for at least one year at 4 °C (expiry date is indicated on the vials).
After reconstitution, the antibodies are stable for several months at 4 °C (if kept sterile) or for minimal one year at -20 °C.
References: Han, G. *et al.* 2005. J. Immunol. 174: 4516-4524
Wang, J. *et al.* 2011. Protein Pept. Lett. 18: 73-83
Wu, F. *et al.* 2009. Vaccine 27: 6095-6101
Yu, H. *et al.* 2007. Virology 359: 264-274

PROUDLY DISTRIBUTED BY

ANIARA
ANIARA DIAGNOSTICA LLC

+1 (513) 770-1991 | +1 (866) 783-3797
7768 Service Center Drive
West Chester, OH 45069
info@aniara.com

For research use only

Version 111031

