


Typical data Human IFN- γ ELISA kit

Catalogue no.: CT201A (5-plate format)

Validated for:

Human



Example of a representative Human IFN- γ standard curve using a 4-parameter logistic fit.

Recommended assay range: 2-128 pg/ml.
Sensitivity: 2 pg/ml.
Specificity: natural and recombinant human IFN- γ .
Calibration: recombinant human IFN- γ .
Type of sample: serum, plasma and culture supernatant.

References list:

- Arend, S.M., Anderson, P., van Meijgaarden, K.E., Skjøt, R.L.V., Subronto, Y.W., van Dissel, J.T. and Ottenhoff, T.H.M. 2000. *J. Infect. Dis.* 181: 1850-1854
- Boer, M.C., van Meijgaarden, K.E., Bastid, J., Ottenhoff, T.H.M. and Joosten, S.A. 2013. *Eur. J. Immunol.* 22: 1925-1932
- Chiriva-Interati, M., Mirandola, L., Yu, Y., Jenkins, M.R., Gornati, R., Bernardini, G., Gioia, M, Chiaramonte, R., Cannon, M.J., Kast, W.M. and Cobos, E. 2011. *J. Immunother.* 34: 490-499
- Geluk, A., van Meijgaarden, K.E., Franken, K.L.M.C., Subronto, Y.W., Wieles, B., Arend, S.M., Sampaio, E.P., de Boer, T., Faber, W.R., Naafs, B. and Ottenhoff, T.H.M. 2002. *Infect. Immun.* 70: 2544-2548
- van den Heuvel, H., Heutinck, M., van der Meer-Prins, E.M.W., Franke-van Dijk, M.E.I., van Miert, P.P.M.C., Zhang, X., Ten Berge, I.J.M. and Claas, F.H.J. 2018. *Hum. Immunol.* 79: 39-50
- Kraaij M.D., Vereyken E.J.F., Leenen P.J.M., van den Bosch T.P.P., Rezaee F., Betjes M.G.H., Baan C.C. and Rowshani A.T. 2014. *Cytokine* 67: 7-12
- Tang, Y-M., Chen, M-H., Chen, G-H., Cai, C-J., He, X-S., Lu, M-G. and Bao, W-M. 2005. *World J. Gastroenterol.* 11: 4574-4578
- Tha-In, T., Metselaar, H.J., Tilanus, H.W., Groothuismink, Z.M.A., Kuipers, E.J., de Man, R.A. and Kwekkeboom, J. 2007. *Blood* 110: 3253-3262

More references can be found in our References Database on our website: www.ucytech.com/references.